



# Patient Warming System Controller Models WC0x Service Manual

**Forward to the Biomedical Engineering Department**  
*For information on operating the HotDog Patient Warming  
System, refer to the "User Manual"*

**Manufactured by:**  
Augustine Temperature Management  
6581 City West Parkway  
Eden Prairie, MN 55344 USA  
TEL 952.465.3500  
FAX 952.465.3501  
[www.hotdogwarming.com](http://www.hotdogwarming.com)

**EU Authorized Representative:**

EC	REP	EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands
----	-----	---

TEL (31) (0) 70 345-8570  
FAX (31) (0) 70 346-7299



## Contents

Introduction.....	3
Indications for Use.....	3
Contraindications.....	3
Warnings.....	3
Caution.....	4
Precautions.....	4
Proper Use and Maintenance.....	4
Read Before Servicing Equipment .....	5
Initial Setup & Assembly.....	5
Contents.....	5
Assembly Procedure .....	5
Mounting the HotDog Controller to an IV Pole .....	6
Overview of Control Panel & Operating Modes .....	8
Alarms .....	9
Maintenance & Testing.....	10
Testing of Indicator Light Function.....	10
Electrical Safety Checks and Functional Testing .....	11
Cleaning .....	13
Cleaning—General .....	13
Cleaning—Controller .....	13
Cleaning—Warming Devices.....	14
Troubleshooting/Error Codes.....	15
Technical Support & Customer Service .....	16
Definition of Symbols.....	17
Accessory Part Numbers.....	17
Specifications .....	18
Electromagnetic Compatibility (EMC).....	20

## INTRODUCTION

The HotDog Patient Warming System consists of the HotDog Controller, reusable warming devices (e.g., Warming Blankets) and accessories. This manual includes maintenance instructions and specifications for the HotDog Controller Models WC0X and V0XX. For information about HotDog warming devices and accessories, refer to the “User Manual” provided with each device/accessory.

The HotDog Controller is designed to help maintain normothermia in patients before, during, and after surgical procedures and to help prevent unintended hypothermia. The system is powered and controlled by an electronic control unit. Warming devices (e.g., blankets) are powered at low voltage, ensuring safety for patients and operators. Warming temperatures are controlled automatically to user-selected levels, and over-temperature safety shut-offs are integrated into the controller as well as into each warming device.

The HotDog Controller can be placed on a flat surface, mounted on an IV pole, or suspended from the OR table/gurney rail using optional hooks. The HotDog Patient Warming System can be operated continuously to maintain uniform heat over the patient. It is the responsibility of the user to determine whether warming is appropriate for each individual patient. The HotDog Patient Warming System should not be used when clinical considerations indicate that warming of the patient is not advisable.

## Indications for Use

The HotDog Patient Warming System is intended to prevent or treat hypothermia and to provide warmth to patients. The HotDog Patient Warming System should be used in circumstances in which patients may not maintain a state of normothermia. The System is intended primarily for use in hospitals and surgical centers including, without limitation, operating, recovery, and emergency rooms and on medical/surgical floors.

## Contraindications

- DO NOT warm ischemic or non-perfused tissue; thermal injury may result. Examples include tissue distal to aortic cross clamping, or when vasoconstrictive drugs would lead to severe, prolonged vasoconstriction.
- DO NOT warm patients receiving transdermal medication; increased drug delivery may occur.

## Warnings

- **WARNING:** To avoid risk of electric shock, this equipment must only be connected to a supply mains with protective earth.

- **EXPLOSION HAZARD – DO NOT** use the HotDog Patient Warming System in the presence of flammable anesthetics or highly oxygen-enriched environments such as hyperbaric chambers, oxygen tents, etc.
- **Inspect HotDog components prior to use** for signs of damage or excessive wear such as cuts, holes, or loose electrical connections. If signs of wear are evident, do not use the product until it is inspected by technical staff.
- **DO NOT** continue to use the HotDog Patient Warming System if the over temperature indicator and/or alarm continue to sound after reset. Refer to the “Alarm” section of this manual for more information.

### **Warming Blanket**

- **DO NOT** place HotDog Warming Blankets under the patient. The Warming Mattress and disposable sheets are the only accessories designed for use under the patient.

### **Accessories and Other Equipment**

- Accessories and cables other than those specified in the User Manual may result in increased emissions or decreased immunity of the HotDog Patient Warming system.
- The HotDog Patient Warming System should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, carefully observe the HotDog Patient Warming System to verify that it operates normally in this non-recommended configuration.

### **Caution**

Federal law (USA) restricts this device to sale by or on the order of a licensed healthcare professional.

### **Precautions**

- Use under the direct supervision of a clinician.
- Monitor the patient’s vital signs regularly during warming according to institutional protocol. If vital sign instability occurs, notify the clinician.
- Care should be taken when using multiple warming methods.
- The risk of skin irritation caused by pooling of surgical prep solutions under the patient may increase with warming; ensure that surgical prep solution instructions for use are followed.

### **Proper Use and Maintenance**

- **DO NOT** continue to use HotDog Warming Blankets beyond the labeled expiration date.

Do not open the HotDog Controller. There are no user serviceable parts. If service is required, contact Technical Support (see **page 16**). The manufacturer assumes no responsibility for the

reliability, performance, or safety of the HotDog Patient Warming System if the following events occur:

- The Controller is disassembled or serviced by an unauthorized person.
- The patient warming system components are used in a manner other than described in the User Manuals.
- The Controller is grounded and should not be attached to an un-grounded table intended for use with a hyfrecator or equivalent devices.

## Read Before Servicing Equipment

Repair, preventive maintenance, safety testing and servicing of the Patient Warming System requires the skill of qualified medical equipment service technicians who are familiar with good practice for medical device repair. Do not open the HotDog Controller. There are no user serviceable parts. If service is required, contact Technical Support (see **page 16**). Perform all maintenance activities in accordance with the instructions in this service manual.

## INITIAL SETUP & ASSEMBLY

### Contents


The following components are included in the HotDog Controller box:

- 1—HotDog Controller (Model WC0X or V0XX)
- 1—IV pole adapter and mounting hardware
- 1—Mains power cord
- 1—CD containing User Manual and Service Manual
- 1—HotDog Warming Blanket Cable (P/N A101)

Reusable HotDog accessories (e.g., Warming Blankets, Warming Mattress, connecting cables, OR table/gurney rail hooks with mounting hardware) and HotDog Disposable Sheets are sold separately.

### Assembly Procedure

1. Remove all items from the box and discard packaging per institutional protocol.
2. Loosen and remove the two cord retainer screws and the cord retainer (see **Figure 1**; the cord retainer is located on the back of the Controller).
3. Firmly insert the mains power cord into the receptacle on the back of the Controller.
4. Place the cord retainer over the strain relief of the mains power cord; replace both screws and tighten to secure the cord retainer (see **Figure 1**).


**Figure 1: Cord Retainer Assembly**

### **Mounting the HotDog Controller to an IV Pole**

To mount the HotDog Controller to an IV pole, place the Controller IV pole adapter around the IV pole and turn the clamp handle clockwise until securely tightened (**Figure 2**). To remove the Controller from the IV pole, turn the clamp handle counterclockwise until the unit releases.


#### **Caution**

To prevent the IV pole from tipping, the Controller must be attached at a height that provides stability. It is recommended to use an IV pole with a minimum wheelbase radius of 35.6 cm (14 in) and to mount the Controller no higher than 112 cm (44 in) from the floor. Failure to properly mount the Controller may result in IV pole tipping, catheter site trauma, and patient injury.

**Figure 2: HotDog Controller Mounted on an IV Pole**

The Controller may also be suspended from the OR table/gurney rail using optional hooks (Figure 3).




**Figure 3: Optional OR Table/Gurney Rail Hooks**



## OVERVIEW OF CONTROL PANEL & OPERATING MODES

Figure 4: HotDog Model WC0X & V0XX Controls

- A—Mains Power ON Light
- B—Temp in Range Light
- C—Temp Adjust Button
- D—Temp Setting Lights (3)
- E—Alarm Port A Light
- F—Alarm Port B Light
- G—Port A (Warming Blanket)
- H—Port B (Auxiliary)
- I—Serial Number
- J—IV Pole Clamp
- K—Power Entry Module
  - Fuse Drawer
  - Mains Power Switch
  - Power Plug
- L—Cord Retainer



**Mains Power Switch / ON Power Indicator**

When the HotDog Controller is plugged into an electrical outlet and the Mains Power Switch on the back of the Controller is turned ON, all displays will illuminate briefly and the Controller will beep. The Mains Power ON Light will illuminate and the Controller will remain idle until a warming device is plugged in. When the unit is ON and idle (i.e., no Temperature Setting Lights are illuminated), no power is applied to the warming device and no alarm conditions are indicated.

**Temperature Adjust Button / Temperature Setting Lights**

Press the Temperature Adjust Button until the desired high (43°C), medium (40°C), or low (37°C) temperature is selected, as indicated by the illuminated Temperature Setting light. The designated warming temperature will be reached within 5 minutes.

**Temp in Range Light**

The green Temperature in Range light will illuminate when the temperature of the Warming Blanket is  $\pm 1^{\circ}\text{C}$  of the selected temperature setting.

**Port A**

Port A should only be used for connection to a HotDog Warming Blanket. When a Warming Blanket is plugged into the Controller, an audible beep indicates that the control and over temperature sensors (safety device) are present and functioning properly and the device is ready for use.

**Port B**

Port B is a 48V output for future approved HotDog accessories.

**Alarms****Alarm Port A**

If the warming device sensor temperature exceeds one degree above set point or other fault conditions exist, an audible alarm sounds and the Alarm Light for the affected port illuminates Yellow. The Controller will automatically turn off power to that warming device. If the Alarm Light stays illuminated and the alarm continues to sound, disconnect the warming device from the Controller to silence the alarm. If the Controller alarms again after a reset was performed, discontinue use and refer the Controller to biomedical engineering for evaluation or contact technical support.

- If Alarm occurs when connecting a Warming Blanket to the Controller, disconnect the Warming Blanket and replace it with another Warming Blanket.

**Alarm Port B**

If the accessory attached to Port B exceeds the pre-set current limit, an audible alarm sounds and the Alarm Port B Light illuminates Yellow. Power is removed from the port. Disconnect the accessory from Port B to reset the alarm.

- If Alarm Port B occurs when connecting an accessory to the Controller, disconnect the accessory and replace it with another accessory.
- If Alarm Port B occurs during use and the Alarm Port B Light illuminates steadily, disconnect the accessory and replace with another accessory.

***Time out timer***

- . The time to reach the set-point temperature from 23 C +/-2 C is less than 10 minutes. If warming device does not reach set temperature within 10 minutes the controller will alarm

**Note:** If the controller is left on for more than 6 hours, the power to the heater is automatically turned off. This is indicated by 3 audible beeps after which the alarm LED will flash continuously. Pushing the temperature select button once will clear this alarm condition and power to the heater will resume.

## **MAINTENANCE & TESTING**

### **Testing of Indicator Light Function**

***Frequency***

This test should be completed upon initial equipment check-in and once every 12 months (or more frequently if required by hospital guidelines).

***Method***

1. Insert the HotDog Controller power plug into a properly grounded hospital grade electrical outlet and confirm that NO cables or devices are connected to any of the ports.
2. Turn the Mains Power Switch to ON and observe for the following proper start-up sequence:
  - a. All lights illuminate briefly.
  - b. The Mains Power ON light remains illuminated while the other lights illuminate one by one in the following order:
    1. Temperature in Range
    2. Alarm A
    3. Alarm B
    4. 43°
    5. 40°
    6. 37°

- c. All of the lights will illuminate simultaneously, and then the unit will emit an audible tone.
  - d. After the sequence completes, only the Mains Power ON Light remains illuminated.
3. If this sequence varies or is incomplete, contact Technical Support (see **page 16**).

## Electrical Safety Checks and Functional Testing

### Frequency

These tests should be completed once every 12 months (or more frequently if required by hospital guidelines).

### Tools/Equipment

- Test fixture (Ref. A115)
- Warming Blanket Cable (Ref. HDPC1, A101 or A102)
- Ground continuity tester
- Leakage current tester
- Calibrated, fast-reacting thermocouple and meter (optional)
- HotDog Warming Blanket (optional)

### Method

1. Perform “Testing of Indicator Light Function” as described on the previous page.
2. Perform the following tests on the Controller per standard institutional protocol:
  - A. Ground continuity
  - B. Connect a Warming Blanket to the Controller and test leakage current to ensure the maximum leakage current does not exceed the requirements in **Table 1**.

**Note: The equipotential stud on the back of the HotDog Controller may be used as a grounding point for these tests.**

Table 1: Maximum Allowable Leakage Current		
Polarity	Condition	Current (mA)
Normal / Reversed	Normal	0.1
	Open Ground	0.5
	Open Neutral	0.5
	Open Ground & Open Neutral	0.5

3. Perform “Functional Testing” described on the following pages.

### **Functional Testing Method for Controller**

Use either a Temperature Test Fixture (Ref. A115) and refer to the included instructions for use, or a HotDog Warming Blanket to perform the steps outlined below. The Test Fixture simulates a Warming Blanket that is operating at 43°C. If a failure is observed during any of these steps, call customer service.


### **Functional Testing for Blanket and Controller**

Use a HotDog Warming Blanket to perform the steps outlined below. If a failure is observed during any of these steps, repeat testing using a different Warming Blanket. If failure is observed with the second Warming Blanket, contact Technical Support (see **page 16**).

1. Tape a calibrated, fast-reacting temperature sensor to the patient-facing surface of the Warming Blanket directly over the sensor marking.
2. Fold the Warming Blanket back on itself (black face to black face) so that the temperature sensor is inside the folded area. Place a 750 to 1000gm weight (such as a small book or notebook) over the sensor location to ensure that the pad remains folded and that there is good contact between the sensor and the folded blanket (

**Figure 5)**

**Figure 5: Warming Blanket Test Configuration**



4. Turn the Mains Power Switch on the Controller to the ON position. Connect the Warming Blanket to the Controller. *The Controller will emit an audible tone when the Blanket is connected.*
5. Set the HotDog Controller to the temperature that is to be verified. If checking all set-points, start with the low temperature setting.

6. After the warming device reaches set point (indicated when the set point readout is no longer flashing), allow the temperature to stabilize for an additional 10 minutes. *NOTE: A temperature overshoot will be noted when testing this way, which is normal.*
7. After 10 minutes, the reading on the temperature meter should be within  $\pm 1^{\circ}\text{C}$  of the set-point temperature. When measuring temperature, the accuracy and tolerance of the meter must be taken into account. This will depend on the type of meter being used and can range from  $\pm 0.2^{\circ}\text{C}$  to  $\pm 2.0^{\circ}\text{C}$ . The measurement tolerance of the meter must be added to the  $\pm 1.0^{\circ}\text{C}$  tolerance for the HotDog system to determine the pass/fail criteria for this test. For example: If the controller is set to  $41^{\circ}\text{C}$  and the measurement is being made with a temperature meter that has a  $\pm 1.0^{\circ}\text{C}$  measurement tolerance, the acceptable range of measured temperatures will be 39 to  $43^{\circ}\text{C}$ . (i.e.  $41 \pm 2^{\circ}\text{C}$ ).
8. Repeat steps 4-6 for the next temperature setting, if required.

## **CLEANING**

### **Cleaning—General**

#### **Warnings**

- DO NOT use a dripping wet cloth and DO NOT immerse HotDog components in liquid. Moisture will damage the components, and thermal injury may result.

#### **Precautions**

- DO NOT use pure harsh solvents (e.g., MEK, acetone, etc.) to clean HotDog components. Solvents may damage plastic parts, labeling and product finish.
- DO NOT use high-level disinfectants (e.g., glutaraldehyde, peracetic acid). The U.S. Centers for Disease Control (CDC) recommends against the use of high level disinfectants for cleaning environmental surfaces that may contact the patient since the chemicals are highly toxic.
- DO NOT spray cleaning solutions into electrical connectors.

#### **Recommended cleaners**

Alcohol-based disinfectants are easiest to use since they are fast-acting and can be either sprayed or wiped on. The following list of alcohol-based cleaners is provided for reference only and is not an endorsement of the manufacturers or their cleaning products: Ecolab (Incidin Liquid, Incides N, Incidin Foam, Incidin Sun, Mikro-Bak III), Merz (Pursept-A Xpress, Pursept Foam, Mucocit-A Economy) and Lysoform (Aerodesin 2000, Lysoform Spray).

Other cleaners that have been tested and are compatible with the outer surfaces of HotDog components include sodium hypochlorite (diluted bleach), phenolic germicidal detergent, and quaternary ammonium detergent.

Iodine containing cleaners may cause discoloration of the surface material and are therefore NOT recommended for routine cleaning. Hydrogen peroxide based cleaning solutions are NOT recommended because the vapors degrade the conductive fabric heaters.

## **Cleaning—Controller**

### ***Frequency***

As needed

### ***Tools/Equipment***

- Sponge or soft cloth
- Mild detergent or anti-microbial spray
- Dry soft cloth

### ***Method***

1. Disconnect the Controller from the power source before cleaning.
2. Wipe unit with moistened sponge or soft cloth; avoid pushing fluids into any openings.
3. Dry with a separate soft cloth.

## **Cleaning—Warming Devices**

### ***Frequency***

Clean between patient use and when the warming device appears soiled.

### ***Method***

Clean the warming devices following protocols for non-critical medical devices that may contact intact skin. Examples of similar devices are blood pressure cuffs, exam table covers, operating room table pads and surgical supports. **Cleaning steps are described in the User Manual provided with the warming device.** Note that the cleaning instructions are general recommendations and are not meant to replace hospital-specific cleaning protocols.

## TROUBLESHOOTING/ERROR CODES

Alarm Mode	Description
Primary over-temp	When the temperature exceeds one degree above nominal (including tolerances), audible and visual alarms are initiated and power is removed from the output. The alarm will reset when: <ul style="list-style-type: none"> <li>• Temperature is within acceptable limits (<math>\pm 1^{\circ}\text{C}</math>)</li> <li>• Cable connecting Warming Blanket to Controller is disconnected</li> <li>• Power is removed at mains</li> </ul>
Secondary over-temp	When the temperature exceeds $46^{\circ}\text{C}$ , audible and visual alarms are initiated. The alarm will reset when: <ul style="list-style-type: none"> <li>• Cable connecting Warming Blanket to Controller is disconnected</li> <li>• Power is removed at mains</li> </ul>
Over-current	When current draw exceeds a predetermined level, audible and visual alarms are initiated. The alarm will reset when: <ul style="list-style-type: none"> <li>• Power is removed at mains</li> </ul>
Failure to reach temp	When the system does not achieve $43^{\circ}\text{C}$ within 10 minutes, audible and visual alarms are initiated. The alarm will reset when: <ul style="list-style-type: none"> <li>• Cable connecting Warming Blanket to Controller is disconnected</li> <li>• Power is removed at mains</li> </ul>
Sensor failure	When either sensor is opened or shorted, audible and visual alarms are initiated and power is removed from output. If both sensors are opened or shorted, no alarms are initiated, power is removed from output, and the temperature selector switch becomes inactive. <ul style="list-style-type: none"> <li>• Replace the Warming Blanket</li> </ul>
Fold Detect	If the Warming Blanket is folded back on itself such that a local over-temperature condition occurs, audible and visual alarms are initiated. The alarm will reset when: <ul style="list-style-type: none"> <li>• Temperature in folded area has cooled, <u>and</u></li> <li>• Cable connecting Warming Blanket to Controller is disconnected, <u>or</u></li> <li>• Power is removed at mains</li> </ul>
Time Out Timer	If warming device does not reach set temperature within 10 minutes the controller will alarm
6-hour time-out timer	If a warming device is left operating for 6 hours with no changes to set point, power will be removed, three short audible chirps will sound, and the alarm Indicators will flash continuously. Pressing the temperature select button will clear the alarm and re-start normal operation.

## **TECHNICAL SUPPORT & CUSTOMER SERVICE**

Please have the serial number of your HotDog Controller when you call for technical support. The serial number is located on the back of the Controller. If it is necessary to return the Controller for service or repair, contact your local supplier or sales representative.

Augustine Temperature Management  
6581 City West Parkway  
Eden Prairie, MN 55344 USA  
TEL 952.465.3500  
FAX 952.465.3501  
[www.hotdogwarming.com](http://www.hotdogwarming.com)

## DEFINITION OF SYMBOLS

	Do Not Place Under Patient		This Side Up		Mains Power On Indicator
	This Side Down		Heating Area		Alarm
	Attention, consult accompanying documents		Reference Number		Lot Number
	BF Patient Applied Part according to IEC60601-1.		Serial Number		Manufacture Date
	Temperature in Range		Transport and storage temperature range		Temperature Adjustment
	Keep Dry		Transport and storage humidity range		Fuse
	Equipotential		EU Authorized Representative		Return to Authorized Representative
	Temperature Sensor		Do not use after YYYY-MM-DD		Manufacturer
<b>IPX2</b>	Protected against dripping water when tilted up to 15°; Vertically dripping water shall have no harmful effect when the enclosure is tilted at an angle up to 15° from its normal position. (The Controller)				
	Conforms to European Medical Device Directive 93/42/EEC				
	Medical Equipment Classified by Intertek Testing Services NA Inc. with respect to electric shock, fire, and mechanical hazards only, in accordance with UL 60601-1. Classified under the Medical Device Directive (93/42/EEC) as a Class IIb device.				

## ACCESSORY PART NUMBERS

The following cable part numbers are used with the HotDog Patient Warming System:

Part Number	Description
A101	HotDog Warming Blankets Cable, 4m (13ft)

## SPECIFICATIONS

### Physical Characteristics

Dimensions	29.21 cm high x 13.97 cm deep x 19.69 cm wide 11.5" high x 5.5" deep x 7.75" wide
Weight	3.6 kg (8.0 lb)
Mounting	Can be clamped to an IV pole

### Temperature Characteristics

Recommended Operating Environment	15°C to 25°C
Temperature Control	Micro-processor
Operating Temperatures	Average temperature at the Warming Blanket:
	High      43° ± 1.0°C      109.4° ± 1.8°F
	Medium    40° ± 1.0°C      104° ± 1.8°F
	Low        37° ± 1.0°C      98.6° ± 1.8°F

### Safety System

Auditory Alarms	Minimum SPL of 65 dB(A) at 3m (from front of controller) with a back ground SPL not to exceed 55dB(A)
Primary Over-temp Alarm	High-Alarm sounds at 44°C + 1°C
Secondary Over-temp Alarm	Independent electronic circuit shuts the heater off if the Warming Blanket temperature reaches 46°C ± 1°C.
Time to reach temperature from 23 C +/-2 C	Less than 10 minutes
Time out timer	If warming device does not reach set temperature within 10 minutes the controller will alarm
Over-current Monitoring	
- Port A	12 amps max.
- Port B	1.7 amps max.
Over-current Protection	Dual input fused lines.

### Electrical Characteristics

Leakage Current	Meets UL 2601-1 and IEC 60601-1 requirements for Class I, Type BF equipment.
-----------------	--

Power Consumption

Peak 580 W

**SPECIFICATIONS (cont'd)**

Power Cord

4.6 m (15 ft.)

Device Ratings

Input: 100-240 VAC, 50/60 Hz, 600VA

Output A: 48 VDC, 500 VA Max

Output B: 48 VDC, 80 VA Max

Fuses

T6.3AL250V (2 x 5x20mm)

Certifications

IEC 60601-1; EN 60601-1-2; UL 60601-1; CAN/CSA-C22.2, No. 601.1, EN 55011



Classification

Classified under IEC 60601-1 Guidelines (and other national versions of the Guidelines) as Class I, Type BF, Ordinary equipment, Continuous operation. Not suitable for use in presence of flammable anesthetic mixtures with air or with oxygen or nitrous oxide. Classified by Intertek Testing Services NA Inc. with respect to electric shock, fire, and mechanical hazards only, in accordance with UL 60601-1. Classified under the Medical Device Directive (93/42/EEC) as a Class IIb device. Classified under Canadian Medical Device Regulation as Class II.

Diagnostics

A qualified technician can perform general system testing. The Controller has no user serviceable parts.

Transport/Storage

Temperature: -20°C to 60°C

Humidity: 20% to 80%

Keep Dry

Important Information

This device complies with the EMC requirements according to IEC 60601-1-2. Radio transmitting equipment, cellular phones, etc. shall not be used in the close proximity of the device since this could influence the performances of the device. Particular precaution must be considered during use of strong emission sources such as High Frequency surgical equipment and similar so that, e.g., the HF-cables are not routed on or near the device. If in doubt, contact a qualified technician or your local representative.

## ELECTROMAGNETIC COMPATIBILITY (EMC)

The HotDog Patient Warming System requires special precautions regarding EMC and must be installed and put into service according to the EMC information provided in this User Manual.

### **Warning**

- **Use of accessories and cables other than those specified may result in increased emissions or decreased immunity of the HotDog Patient Warming System.**
- **The HotDog Patient Warming System should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, carefully observe the HotDog Patient Warming System to verify that it operates normally in this configuration.**

Guidance and Manufacturer's Declaration – Electromagnetic Emissions		
The HotDog™ Patient Warming System is intended for use in the electromagnetic environment specified below. The customer or user of the HotDog Patient Warming System should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic Environment – Guidance
RF emissions, CISPR 11	Group 1	The HotDog Patient Warming System uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions, CISPR 11	Class A	The HotDog Patient Warming System is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions, IEC 61000-3-2	Class A	
Voltage fluctuations/ flicker emissions, IEC 61000-3-3	Complies	

<b>Guidance and Manufacturer's Declaration – Electromagnetic Immunity</b>			
The HotDog™ Patient Warming System is intended for use in the electromagnetic environment specified below. The customer or the user of the HotDog Patient Warming System should assure that it is used in such an environment.			
<b>Immunity Test</b>	<b>IEC 60601 Test Level</b>	<b>Compliance Level</b>	<b>Electromagnetic Environment – Guidance</b>
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line(s) to line(s) ±2 kV line(s) to earth	±1 kV line(s) to line(s) ±2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % <i>UT</i> (>95 % dip in <i>UT</i> ) for 0,5 cycle 40 % <i>UT</i> (60 % dip in <i>UT</i> ) for 5 cycles 70 % <i>UT</i> (30 % dip in <i>UT</i> ) for 25 cycles <5 % <i>UT</i> (>95 % dip in <i>UT</i> ) for 5 sec	<5 % <i>UT</i> (>95 % dip in <i>UT</i> ) for 0,5 cycle 40 % <i>UT</i> (60 % dip in <i>UT</i> ) for 5 cycles 70 % <i>UT</i> (30 % dip in <i>UT</i> ) for 25 cycles <5 % <i>UT</i> (>95 % dip in <i>UT</i> ) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the HotDog Patient Warming System requires continued operation during power mains interruptions, it is recommended that the HotDog Patient Warming System be powered from an uninterruptible power supply or a battery.
Power frequency 50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE <i>UT</i> is the a.c. mains voltage prior to application of the test level.			

Guidance and Manufacturer’s Declaration – Electromagnetic Immunity (cont’d)			
The HotDog™ Patient Warming System is intended for use in the electromagnetic environment specified below. The customer or the user of the HotDog Patient Warming System should assure that it is used in such an environment.			
Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment – Guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 V	<p>Portable and mobile RF communications equipment should be used no closer to any part of the HotDog Patient Warming System, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p><b>Recommended separation distance</b></p> $d = 1,2\sqrt{P}$ $d = 0,35\sqrt{P} \quad 80 \text{ MHz to } 800 \text{ MHz}$ $d = 0,7\sqrt{P} \quad 800 \text{ MHz to } 2,5 \text{ GHz}$ <p>where <math>P</math> is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and <math>d</math> is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,<sup>a</sup> should be less than the compliance level in each frequency range.<sup>b</sup></p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> 
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2,5 GHz	10 V/m	
<p>NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.</p> <p>NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>			
<p><sup>a</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the HotDog Patient Warming System is used exceeds the applicable RF compliance level above, the HotDog Patient Warming System should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the HotDog Patient Warming System.</p> <p><sup>b</sup> Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.</p>			

**Recommended separation distances between portable and mobile RF communications equipment and the HotDog Patient Warming System**

The HotDog™ Patient Warming System is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the HotDog Patient Warming System can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the HotDog Patient Warming System as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = 1,2\sqrt{P}$	80 MHz to 800 MHz $d = 0,35\sqrt{P}$	800 MHz to 2,5 GHz $d = 0,7\sqrt{P}$
0,01	0,12	0,04	0,07
0,1	0,37	0,11	0,22
1	1,2	0,35	0,70
10	3,7	1,1	2,2
100	12	3,5	7,0

For transmitters rated at a maximum output power not listed above, the recommended separation distance  $d$  in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where  $P$  is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.